ALL INDIA INSTITUTE OF MEDICAL SCIENCES, DEOGHAR अखिल भारतीय आयुर्विज्ञान संस्थान, देवघर

NSTITUTE OF NATIONAL IMPORTANCE UNDER MINISTRY OF HEALTH AND FAMILY WELFARE (स्वास्थ्य और परिवार कल्याण मंत्रालय के अंतर्गत राष्ट्रीय महत्व की संस्थान)

P.T.I., Daburgram, Jasidih, Deoghar, (Jharkhand) - 814142 पी.टी.आई., डाबरग्राम, जसीडीह, देवघर, (झारखंड) - ८१४९४२

आरोग्यम् परमं सुखम्

Invitation of quotation for Equipment list for Pediatric OPD at AIIMS Deoghar

Reference No.: AIIMS/Deoghar/Pediatric OPD / Equipment list /2021-22/40

Date of Issue: 02nd February, 2021

Last Date of Submission: 12th February, 2021 at 04:00 PM.

All India Institute of Medical Sciences, Deoghar

P.T.I. campus, Daburgram, Jasidih, Deoghar: 814142, Jharkhand

Email: office.aiimsdeoghar@gmail.com

Invitation of quotation for Equipment list for Pediatric OPD at AIIMS Deoghar.

Sealed Quotations are invited on behalf of AIIMS, Deoghar for Equipment list for Pediatric OPD *(Annexure-I)* required for the Institute as per terms & conditions mentioned below. The filled quotations along with the entire required document must reach in the office of the undersigned on or before **12.02.2021** at **04:00 PM**. The **Envelope containing the quotation** must be sealed and **super scribed as under:**-

"QUOTATION FOR MINOR PROCEDURE STANDING LIGHT FOR DEPT. OF ENT AT AIIMS DEOGHAR FOR REFENRENCE NO.: AIIMS/DEOGHAR/ PEDIATRIC OPD / EQUIPMENT LIST /2021-22/40 **DUE ON 12.02.2021 at 04.00 PM"**

The Quotation should be send to the address:-

Assistant Procurement Officer AIIMS Deoghar PTI Campus, Daburgram, Jasidih, Deoghar, Jharkhand -814142.

1. Terms & Conditions:

- a) The quotations received **after this deadline or unsealed shall not be entertained** under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. The offer submitted by Fax / email shall not be considered and no correspondence will be entertained in this matter.
- b) Quotations must be in the **enclosed prescribed Performa** (Annexure -2) on the **letter head of the firm duly signed** by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- c) The supplier should supply all the equipments / items along with its accessories.
- d) Rates must be quoted in Indian rupees.
- e) Rates must be inclusive of all charges (including Freight charges, Insurance, installation, taxes etc.).
- f) **No overwriting or cutting** is permitted in the rate. If found, the quotation shall be summarily rejected.
- g) The rates quoted must be valid for 90 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- h) **Total cost/amount will be taken in consideration for L1.** Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- i) The firm / agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - The firm shall have valid GST / Other taxes and IT PAN.
 - The firm should not be black listed by any Government agency/Department.
 - Purchase order of any government institute for similar work.
- j) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers

and it will be rejected accordingly. Any conditional quotation shall be rejected summarily.

- k) **Delivery Period** within 15 days from Purchase order.
- l) **Liquidated Damage**: If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- m) **Payment Terms**: Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Deoghar.
- n) 5% amount from bill will be deducted as security deposit. It will be refunded within three months after satisfactory performance report submitted by the nominated person/concern Department, for the equipments/work.
- o) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Deoghar with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to arbitrator appointed by The Executive Director, AIIMS, Deoghar, whose decision will be final and binding upon the contractor.
- p) AIIMS, Deoghar reserves the right to increase or decrease quantity. Decision of Quantity of items in the AIIMS, Deoghar will be final in this regard.
- q) AIIMS, Deoghar reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of The Executive Director, AIIMS, Deoghar, will be final in this regard.
- r) The envelope containing quotation should be sealed with WAX/TAPE on both sides. The quotation received after due date will not be accepted.

Encl.: Annexure 1 (Specification)

Annexure 2 (Format of price bid)

Reference no: AIIMS/Deoghar/Pediatric OPD / Equipment list /2021-22/40

S.	Name of instruments	Specifications	Quantity
S. no. 1.	Name of instruments Digital Weighing machine (neonate and infant) -	a) microprocessor based digital weighing scale with facility to weigh lying down. b)weight range 0-20 kg, resolution 5 gms c)minimum weight measurement 20 gms d) accuracy +/-5gms,resolution 5 gms e)weighing unit –standard display in gms should have kg/lbs switch over f)functions -tare -automatic switch off -auto hold g)should have tare range:upto 10 kg h)automatic switchoff after 5-10 minutes of nonuse i)should have inbuilt handle as well as carrying case to make the battery/powered scale easy to transport anywhere to use j)Bassinet-curved surface to prevent baby to fall	Quantity 02 pc
		k)Bassinet size-600*300mm+/-25mm l)made up of high quality sturdy plastic/acrylic tray;easy to clean m)LED display enough to be visible from distance of 3-4 feet to normal eye n)the unit should be able to run on mains and /or in built battery with power supply of 220/240 v,50-60hz with DC adapter o)should be USA FDA and CE approved(certificate to be submitted) p)ISO 9001 certified manufacturer(certificate to be submitted q)training and installation at end user site r)availability of spares for atleast 7 years after date of	
		installation s)item covered under warranty t)supplied with -1 spares set of fuses -user manual with trouble shooting guidance -technical manual with maintenance and first line technical intervention instruction in English -list of price of spare parts -rates of consumable should be quoted separately	
2.	Infantometer	 a) Firm and Flat Horizontal Surface, b) Graduation for Infantometer (cm)-0.1 cm, c) Accuracy for Infantometer (cm):+/-0.1cm, d) Precision for Infantometer (cm):+/-0.1cm, e) Width of board for infantometer in cm-Approx 30 cm, f) Length of Infantometer For Measuring in cm-100 centimeter, Weight of Infantometer (in kg) 1.5 kilogram, g) Material of board-ABS Material make 	02 pc

	I	/ A ample mitable in the discount of the Company			
		(Acrylonitrile butadiene styrene), Thickness of Material of Board-6 millimeter, Fold Up			
		mechanism,			
		h) Double sided Graduation, Movable Foot Piece			
		Perpendicular to the measuring Board,			
		Antiskid.			
3.	Stadiometer	a) Confirmity to Stadiometer as per ICDS	02 pc		
		standards measuring range 20-205 cm			
		minimum, Smallest Graduation for			
		Stadiometer (cm)-0.1 cm,			
		b) Accuracy for Stadiometer (cm)-(+/-)0.2 cm,			
		c) Precision for Stadiometer (cm)-(+/-)0.2 cm,d) Material of Board-Aluminium.			
4.	Blood pressure	d) Material of Board-Aluminium.a) Aneroid Type Of SPHYGMOMANOMETERS or	02 pc		
4.	apparatus aneroid type	Blood Pressure recording unit, Shock proof	02 μς		
		b) Latex-Free			
		c) Exclusive Pressure-Distribution System			
		d) Pressure-Proof Manometer. Ensures durability			
		and reliability			
		e) Micro-Filter			
		f) Protects valve and manometer			
		g) Large-diameter Scale			
		h) 56 mm diameter black-on-white scale is easy			
		to read i) High Quality Indicator. Precisely calibrated,			
		durable and shock resistant			
		j) Large Insufflation Bulb. 50% larger bulb			
		volume for fast inflation of the cuff			
		k) Exclusive Air-Release Valve			
		I) Provides fine and precise air release rate			
		adjustment, quick deflation, and extreme			
		durability			
		m) A rubber-stop prevents valve-lock. Upper			
		housing made of aluminum alloy with a shock absorbing ring protecting indicator and glass			
		n) Lower housing is moulded of shockproof			
		plastic with shock absorbing, anti-slip coating.			
		Light alloy, adjustable Spoon.			
		o) For both right- or left-handed operation			
		p) High quality, single-tube cuff with Velcro			
		Fastener			
		q) Washable			
		r) Portable, Measurement Method-Oscillometric			
		s) Range: Pressure measurement (mmHg)-0 to 300			
		t) Accuracy: Pressure measurement +/- 3 mm hg			
5.	Peak flow meter	a)The device is a manual, hand-held instrument	02 pc		
		typically consists of a tube for patient exhalation, an	•		
		easy-to-grip handle, and a calibrated scale that shows			
		the value of the peak flow			
		b)Main body size without mouth piece (LxWxT) in mm 90X50X20			
		c)Material of hand piece or main body Impact			
		resistant polycarbonate			
		d)Weight of the instrument (gms) 60 to 70			
		e)Material of mouth piece High density			
		polyethylene			

		f)Usage of mouth piece Single use g)h)Mouth piece length (mm) 64 h)Mouth piece outer dia (mm) 22 i)Scale used on chart ERS (European Respiratory Society),ATS (American Thoracic Society) j)Parts of the instrument come in contact with patient are cleanable with dry cloth impregnated in 70% alcohol Yes PERFORMANCE PARAMETERS k)Measurement range(Lts/min) 60 to 800 l)Accuracy as per ISO 23747:2007 shall be ± 10% of measurement m)Scale resolution 10 LPM for range 30 to 700 and 20 LPM for range 700 to 800 n)Maximum Noise level (dB) 150 o)Resetting value for next use is simple and easy for patients with limited dexterity p)Reproducibility is +/- 5% of measurement Yes CERTIFICATIONS & REPORTS q)Submission of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification at the time of supply r)Product certification ISO: 23747-2007 s)Four digit number of notified body if the product is EU-CE certified t)Manufacturer's company Certification to ISO is available u)Seller should submit all necessary certificates, test reports and licenses related to the product at the time of supply which may be verified by the buyer v)WARRANTY & MAINTENANCE w)Warranty (Including spares and calibration) (Yrs) 3 year	03
6.	Ambu bag and Face mask - Neonate & infant		02 sets

7.	Ambu bag & Face mask – Children	Autoclavable, Manual resuscitator or self-inflating bag with reservoir bag, Low resistance Non-rebreathing expiratory valve available, Pressure limiting valve available Face mask size- 1,2,3,4,5(round shaped and anatomical) DIMENSIONS Maximum bag Volume: 500-1000 ml,1000-1500 ml, 1500-2500 ml. MATERIAL Material (Medical grade, latex-free, non-toxic, non-allergic and antistatic), Silicon rubber FUNCTIONAL a)Application temperature 20 deg C to 50 deg C b)Storage Temperature - 40 deg C to 70 deg C c)Ports available in the bag to connect with Oxygen	02 sets
8.	Laryngoscope - neonate Infant	and reservoir a) Standard Laryngoscope, Shape of Blade-Curved and straight, Size of Blades -00,0,1,2, Number of blades included- one set of each type straight and curved blade b) Type of Handle-infant size c) Handle with Universal Adapter for Interchangeable blades, d) Light Source-LED bulb, e) Fitting of light source-Bulb on the blade, f) Material of handle of Laryngoscope-Stainless Steel Medical grade, g) Provision for locking or unlocking the blade, Type of Handles-Hook-on fitting., child Multilingual user's instruction manual in protective case. Additional -Should be stainless steel matt finishedShould provide straight blades - 2Nos each -Should provide spare bulb - 6 no	02 sets
9.	Laryngoscope – Children	-Should provide spare bulb — 6 no a) Standard Laryngoscope, b) Shape of Blade-Curved and straight, c) Size of Blades -2,3,4 of both straight and curved, d) Number of blades included- one set of each type of blade e) Type of Handle-paediatric f) Handle with Universal Adapter for Interchangeable blades, g) Light Source-LED bulb, h) Fitting of light source-Bulb on the blade, i) Material of handle of Laryngoscope-Stainless Steel Medical grade, j)Provision for locking or unlocking the blade, Type of Handles-Hook-on fitting, child Multilingual user's instruction manual in protective case. k) should provide battery and spare bulbs	02 sets

10.	BP cuffs	2 sets of different sizes cuff new born (4cm X8cm), Infant (6cmX12cm) Children (9cm X 18cm) Adult (12cmX22cm)	02 sets
11.	Pediatric Electrocardiography EKG Patch Electrodes	-No Effect of high temprature during summer -Electrodes should have good sticking material, Solid Gel covered by Translucent or Transparent Thin Sheet. -Hypoallergic certificate should be attached -Brand mark on the White lebel -Suitable for hairy chest -Stainless Steel Buttons Metal snap for attachment to ECG Electrodes Certification: ISO /USFDAcertified Disposable ECG Electrodes type Pediatric Sensor material Silver	200

Reference no:	Date:
---------------	-------

[Letter head of firm]

	<u>PR</u>	ICE BII	O FORM	_		
AIIM	stant Procurement Officer, IS, Deoghar. khand					
Dear	· Sir,					
EQU REF	eam/are submitt IPMENT LIST FOR DEPARTMENT ENRENCE NO.: AIIMS/DEOGHAR/ P ON 12.02.2021 at 04.00 PM" for De	OF P	EDIATRICS RIC OPD	S OPD AT AI / EQUIPMEN	IIMS DEOGH Γ LIST /202	HAR FOR
	 I/We have thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly. I/We hereby offer to supply at the following rates: 					
S. no.	Name of Item with specification	Qty	Quoted make	Unit price	GST	Total cost
1						
2						
	Total cost/ amount					
Tota	l cost/amount (in words)	1				
Note	x-					
 The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED. Catalog must be attached with quotation for technical evaluation. 						
	Date:					
	(Name):					
	Place:					
	Name of Firm/Company/Agency:					
	GSTIN No.: :					
	Phone No:					
	Email:					
	(Signature of Authorized Person)					